


# AURELIA® Bold®

Examination Gloves



<b>73995</b>	(XS)	ø 100	10 x 100
<b>73996</b>	(S)	ø 100	10 x 100
<b>73997</b>	(M)	ø 100	10 x 100
<b>73998</b>	(L)	ø 100	10 x 100
<b>73999</b>	(XL)	ø 100	10 x 100



**Nitrile**



**Powder Free**



**Black**



**5.5 +/- 0.2 g**



**Micro-Textured**




**5 Years**

Premium quality Black Nitrile examination gloves that offer greater tensile strength and protection than both Latex and Vinyl.

Nitrile examination gloves offer extra comfort and flexibility.

- 5mil Powder-Free Nitrile
- Non-Sterile
- Beaded cuff
- Ambidextrous
- Latex Free
- AQL 1.5

## Sizes & Dimensions



	Width (mm)	Length (mm)
<b>XS</b>	70 ± 10 mm	min. 240 mm
<b>S</b>	80 ± 10 mm	min. 240 mm
<b>M</b>	95 ± 10 mm	min. 240 mm
<b>L</b>	110 ± 10 mm	min. 240 mm
<b>XL</b>	≥ 110 mm	min. 240 mm

## Single wall thickness



- ① min. 0.13 mm
- ② min. 0.10 mm
- ③ min. 0.08 mm

## Physical properties

	Unaged	Aged
Tensile Strength	min. 14.0 MPa	min. 14.0 MPa
Elongation	min. 500 %	min. 400 %
Strength at Break	min. 6N	min. 6N

Aurelia Gloves conforms and complies with:

- The general safety and performance requirements of FDA Medical Device Regulation for Class 1 medical devices (21 CFR 880.6250) and complies with all general controls (section 513(a)(1)(A) of the Federal Food, Drug, and Cosmetic Act (Act); 21 U.S.C. 360c(a)(1)(A))
- The standard specification requirements for ASTM D6319- Freedom From Holes, Physical Dimensions Test, Physical Requirement Test and Packaging
- The standard specification requirements for ASTM D6978-05 - Standard Practice for Assessment of Resistance of Medical Gloves to Permeation by Chemotherapy Drugs
- EEC regulations concerning the conformity of materials and products that are allowed to come into contact with food. In accordance with Regulation EEC 1935/2004, Regulation EC 10/2011 & Regulation (EC) No 2023/2006.
- The gloves are certified manufactured according to ISO 9001:2015 and ISO 13485:2016 Quality Management Systems