

# AURELIA® Ignite®

Examination Gloves



<b>97887</b>	(M)	ø 100	10 x 100
<b>97888</b>	(L)	ø 100	10 x 100
<b>97889</b>	(XL)	ø 90	10 x 90
<b>97880</b>	(XXL)	ø 90	10 x 90



**Nitrile**



**Powder Free**



**Orange**



**8.4 +/- 0.2 g**



**Diamond-Textured**




**5 Years**

Premium quality heavy-duty Orange Nitrile examination gloves with an excellent diamond grip finish which offers maximum grip, increased tensile strength, optimum protection in tough environments and superior tactile sensitivity.

- 7mil Powder-Free Nitrile
- Non-Sterile
- Beaded cuff
- Ambidextrous
- Latex Free
- AQL 1.5

## Sizes & Dimensions



M	95 ± 10 mm	min. 240 mm
L	110 ± 10 mm	min. 240 mm
XL	120 ± 10 mm	min. 240 mm
XXL	≥ 120 mm	min. 240 mm

## Single wall thickness



- ① min. 0.25 mm
- ② min. 0.20 mm
- ③ min. 0.09 mm

## Physical properties

	Unaged	Aged
Tensile Strength	min. 14.0 MPa	min. 14.0 MPa
Elongation	min. 500 %	min. 400 %
Strength at Break	min. 6N	min. 6N

Aurelia Gloves conforms and complies with:

- The general safety and performance requirements of FDA Medical Device Regulation for Class 1 medical devices (21 CFR 880.6250) and complies with all general controls (section 513(a)(1)(A) of the Federal Food, Drug, and Cosmetic Act (Act); 21 U.S.C. 360c(a)(1)(A))
- The standard specification requirements for ASTM D6319- Freedom From Holes, Physical Dimensions Test, Physical Requirement Test and Packaging
- The standard specification requirements for ASTM D6978-05 - Standard Practice for Assessment of Resistance of Medical Gloves to Permeation by Chemotherapy Drugs
- EEC regulations concerning the conformity of materials and products that are allowed to come into contact with food. In accordance with Regulation EEC 1935/2004, Regulation EC 10/2011 & Regulation (EC) No 2023/2006.
- The gloves are certified manufactured according to ISO 9001:2015 and ISO 13485:2016 Quality Management Systems