



AURELIA® Sonic®

Examination Gloves



9377B5 (XS)	ø 1300	10 x 100
9377B6 (S)	ø 1300	10 x 100
9377B7 (M)	ø 1300	10 x 100
9377B8 (L)	ø 1300	10 x 100
9377B9 (XL)	ø 1300	10 x 100



Nitrile



Powder Free



Cobalt Blue



3.2+/- 0.2 g



Finger-Textured




5 Years

Quality Cobalt Blue Nitrile examination gloves that offer increased tensile strength and protection over Latex and Vinyl.

Incredible Ultra-Thin Nitrile examination gloves specifically designed to offer increased comfort, flexibility & sensitivity.

- 2.2mil Powder-Free Nitrile
- Non-Sterile
- Beaded cuff
- Ambidextrous
- Latex Free
- AQL 1.5

Sizes & Dimensions



XS	70 ± 10 mm	min. 240 mm
S	80 ± 10 mm	min. 240 mm
M	95 ± 10 mm	min. 240 mm
L	110 ± 10 mm	min. 240 mm
XL	≥ 110 mm	min. 240 mm

Single wall thickness



- ① min. 0.11 mm
- ② min. 0.05 mm
- ③ min. 0.05 mm

Physical properties

	Unaged	Aged
Tensile Strength	min. 14.0 MPa	min. 14.0 MPa
Elongation	min. 500 %	min. 400 %
Strength at Break	min. 6N	min. 6N

Aurelia Gloves conforms and complies with:

- The general safety and performance requirements of FDA Medical Device Regulation for Class 1 medical devices (21 CFR 880.6250) and complies with all general controls (section 513(a)(1)(A) of the Federal Food, Drug, and Cosmetic Act (Act); 21 U.S.C. 360c(a)(1)(A))
- The standard specification requirements for ASTM D6319- Freedom From Holes, Physical Dimensions Test, Physical Requirement Test and Packaging
- The standard specification requirements for ASTM D6978-05 - Standard Practice for Assessment of Resistance of Medical Gloves to Permeation by Chemotherapy Drugs
- EEC regulations concerning the conformity of materials and products that are allowed to come into contact with food. In accordance with Regulation EEC 1935/2004, Regulation EC 10/2011 & Regulation (EC) No 2023/2006.
- The gloves are certified manufactured according to ISO 9001:2015 and ISO 13485:2016 Quality Management Systems