

**TECHNICAL DATA SHEET**
**Polychloroprene, Non-Sterile Medical Examination Glove: ASTM D 6977-04**
**Intended Use**

Non-sterile, single use, medical examination gloves are intended to prevent transmission of a wide variety of diseases to both patients and health care personnel. Specifically, gloves are intended to act as a barrier between the healthcare provider and patients to prevent contamination between patient and caregiver.

ASTM D 6977-04						
Characteristic			Testing Methodology	Acceptable Limit		
Dimension & Tolerance	Width (mm)		ASTM D 3767	Extra-Small (XS)	72 – 80	
				Small (S)	82 – 90	
				Medium (M)	91 – 99	
				Large (L)	102 – 110	
				X-Large (XL)	111 – 119	
	Length (mm)				≥ 240	
	Thickness (mm)	Palm			0.08 – 0.12	
Finger			0.11 – 0.15			
Cuff			0.06 – 0.10			
Physical Requirements	Before ageing	Tensile Strength	ASTM D 412	≥ 14 MPa		
		Ultimate Elongation		≥ 500%		
	After ageing	Tensile Strength		≥ 14 MPa		
		Ultimate Elongation		≥ 400%		
Powder Residue	Powder Free Gloves		ASTM D 6124	≤ 2,0 mg/glove, Inspection Level N=5		
Description	Ambidextrous					
Color	Emerald					



#### Additional Information

Storage Condition	Country of Origin
Store in dry, cool, and ventilated area. Avoid excessive heat (40°C, 104°F). Shield open box from direct sun and fluorescent lighting and X-rays.	Indonesia

#### Appendix 1 List of Item Codes and Attributes

## Appendix 1 (List of items)

AMD Medicom Inc.

Item Code	Description	Packaging Information	Size	Color	Specific Attribute	Treatment	Marketed in USA as	Marketed in Canada as
1228	Medicom Ocean Pacific NeoNatural Powder-Free Textured Chloroprene Medical Examination Gloves	100 Gloves/Box, 10 Boxes/Case, 1000 Gloves/Case	A = X-Small, B = Small, C = Medium, D = Large, E = X-Large	Emerald	Country of origin: Indonesia	Non-Sterile	Medical Device Class 1	Medical Device Class 2